



America

CERTIFICATE

No. U8 17 05 42035 007

Holder of Certificate: **INTEGRA Biosciences AG**

Tardisstrasse 201
7205 Zizers
SWITZERLAND

Production Facility(ies): 42035

Certification Mark:



C US

Product: **Laboratory Equipment**

Model(s): **VIAFLO 96; VIAFLO 384**

Parameters:

Rated voltage:	100-240 V AC
Rated frequency:	50/60 Hz
Rated power:	400 W
Protection class:	I

Tested according to:

UL 61010-1:2012/R:2015-07
CAN/CSA-C22.2 No. 61010-1:2012 + UPD No. 1:2015-07 + UPD No. 2:2016-04
UL 61010-2-081:2015-08
CAN/CSA-C22.2 No. 61010-2-081:2015-11

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 028-713105220-000

Date, 2017-05-23

Page 1 of 1

