

Channel	Part No.	Volume Range (µl)	Volume Increments (µl)	Test Volume (µl)	Accuracy (±%)	Precision (≤%)
1	3011	0.2 – 2.0	0.001	0	12.00	6.00
				1	3.00	1.60
				2	1.80	1.00
	3012	1 – 10	0.01	1	5.00	3.00
				5	1.50	0.80
				10	1.20	0.50
	3013	2 – 20	0.01	2	7.00	2.00
				10	1.60	0.50
20				1.20	0.35	
3015	10 – 100	0.1	10	3.50	1.00	
			50	1.00	0.35	
			100	0.80	0.20	
3016	20 – 200	0.1	20	2.50	1.00	
			100	1.00	0.30	
			200	0.70	0.18	
3018	100 – 1000	1	100	3.00	0.60	
			500	1.00	0.30	
			1000	0.80	0.20	
3019	500 – 5000	5	500	3.00	0.70	
			2500	1.20	0.30	
			5000	0.60	0.18	
8	3022	1 – 10	0.01	1	12.00	7.00
				5	4.00	3.00
				10	2.00	1.20
	3024	5 – 50	0.05	5	5.00	2.00
25				2.00	1.00	
50				1.20	0.50	
3025	10 – 100	0.5	10	5.00	2.00	
			50	2.00	0.80	
			100	1.20	0.50	
3026	20 – 200	0.5	20	5.00	1.30	
			100	2.00	0.50	
			200	1.20	0.40	
12	3032	1 – 10	0.01	1	12.00	7.00
				5	4.00	3.00
				10	2.00	1.20
	3034	5 – 50	0.05	5	5.00	2.00
25				2.00	1.00	
50				1.20	0.50	
3035	10 – 100	0.5	10	5.00	2.00	
			50	2.00	0.80	
			100	1.20	0.50	
3036	20 – 200	0.5	20	5.00	1.30	
			100	2.00	0.50	
			200	1.20	0.40	

1. The multichannel specifications represent the performance of all channels of a pipette.

2. Precision = coefficient of variation