

# Certificate of conformance

# INTEGRA

**Product name:** 25 ml Reservoir      **Product number:** 4311

**Lot Number:** 99020654, Endotoxin-Free

## Our commitment to customers:

**INTEGRA** polymer products are using precise and consistent manufacturing techniques that ensure quality and excellent performance. We strive to exceed our customers' expectations and encourage continuous feedback. The reservoirs are manufactured in accordance to ISO 9001.

Our product lots are traceable down to the corresponding raw materials used and date produced.

Our manufacturing process incorporates consistent and continuous inspection throughout all phases of the manufacturing process. The reservoirs produced under these strict conditions are contaminant free and consistent from lot-to-lot.

## High Sensitivity Endotoxin Testing:

For products used in endotoxin testing our Swiss reference lab, accredited to ISO 17025, performed a lot-specific Limulus amoebocyte lysate (LAL) test using the turbidimetric test procedure according to EP 2.6.14 / USP <85>. The endotoxin acceptance level for this certificate is < 0.005 EU/ml.

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.

Email: [info@Integra-biosciences.com](mailto:info@Integra-biosciences.com)

[www.integra-biosciences.com](http://www.integra-biosciences.com)

QUALITY-CERT\_Endotx\_V00

