



# Declaration of conformity

INTEGRA Biosciences AG – 7205 Zizers, Switzerland

declares on its own responsibility that the devices

Description	Models
<b>FIREBOY plus</b>	<b>144000</b>
<b>FIREBOY eco</b>	<b>144010</b>

comply with:

## EU Directives

Low Voltage Equipment	<b>2014/35/EU</b>
Electromagnetic Compatibility	<b>2014/30/EU</b>
Restriction of Hazardous Substances	<b>2011/65/EU</b>
Waste Electrical and Electronic Equipment	<b>2012/19/EU</b>
Battery Directive	<b>2006/66/EC</b>

## EU Regulations

Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH)	<b>1907/2006</b>
Capacity Labelling of Portable Secondary Batteries	<b>1103/2010</b>
Ecodesign - Power supplies	<b>278/2009</b>

## Standards for EU


Safety requirements for electrical equipment for measurement, control and laboratory use - General requirements.	<b>EN 61010-1: 2010</b>
Electrical equipment for measurement, control and laboratory use - EMC requirements.	<b>EN 61326-1: 2013</b>

## Standards for Canada and USA

Safety requirements for electrical equipment for measurement, control and laboratory use - General requirements.	<b>CAN/CSA-C22.2 No. 61010-1</b>
Safety requirements for electrical equipment for measurement, control and laboratory use - General requirements.	<b>UL 61010-1</b>
Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.	<b>Part 15 of the FCC Rules Class A</b>

Zizers, March 02, 2020

  
Urs Hartmann  
CEO

  
Thomas Neher  
Quality Manager