



## Declaration of Conformity

**INTEGRA Biosciences AG – 7205 Zizers, Switzerland**

declares on its own responsibility that the devices

| Description                      | Models        |
|----------------------------------|---------------|
| <b>WELLJET Dispenser</b>         | <b>5000</b>   |
| <b>WELLJET Dispenser Stacker</b> | <b>5001</b>   |
| <b>Mains adapter</b>             | <b>128909</b> |

comply with:

| <b>EU Provisions</b> | <b>Scope</b>   |
|----------------------|--|
| 2014/35/EU           | Low voltage directive (LVD)  |
| 2014/30/EU           | Electromagnetic compatibility (EMC)  |
| 2014/53/EU           | Radio Equipment directive (RED)  |
| 2012/19/EU           | Waste electrical and electronic equipment (WEEE)                             |
| 2011/65/EU           | Restriction of hazardous substances (RoHS)                                   |
| 1907/2006            | Registration, evaluation, authorisation and restriction of chemicals (REACH) |
| EN 9001:2015         | Quality Management   |
| EN 61010-1:2020      | Safety general laboratory equipment  |
| EN 61326-1:2013      | Electromagnetic compatibility laboratory equipment                           |
| EN 61010-2-081:2020  | Safety automatic laboratory equipment  |
| EN 62368-1:2021      | Safety information technology equipment                                      |
| EN 301 489-1 V2.2.3  | EMC radio equipment: technical requirements                                  |
| EN 301 489-3 V2.2.0  | EMC radio equipment: conditions for operating                                |
| EN 300 330 V2.1.1    | Short range devices  |
| EN 50364:2019        | Human exposure to electromagnetic fields                                     |

## WELLJET – Declaration of conformity

---

| <b>UK Provisions</b> | <b>Scope</b>                                     |
|----------------------|--|
| S.I. 2016/1101       | Electrical equipment safety                      |
| S.I. 2016/1091       | Electromagnetic compatibility (EMC)              |
| S.I. 2017/1206       | Radio equipment (RED)                            |
| S.I. 2013/3113       | Waste electrical and electronic equipment (WEEE) |
| S.I. 2012/3032       | Restriction of hazardous substances (RoHS)       |
| BS 61010-1:2010      | Safety general laboratory equipment              |
| BS 61010-2-081:2020  | Safety automatic laboratory equipment            |
| BS 62368-1:2020      | Safety information technology equipment          |
| BS 63000:2018        | Restriction of hazardous substances (RoHS)       |

| <b>USA Provisions</b>    | <b>Scope</b>  |
|--------------------------|---|
| 47 CFR Part 15 (FCC)     | Electromagnetic compatibility (EMC)                               |
| 17 CFR Parts 240 & 249b  | Dodd frank “Conflict minerals”                                    |
| 27 CCR Parts 25102-27001 | Proposition 65: The safe drinking water and toxic enforcement act |
| TSCA 40 CFR Part 751     | Toxic substances control act                                      |
| UL 61010-1:2012          | Safety general laboratory equipment                               |
| UL 61010-2-081:2019      | Safety automatic laboratory equipment                             |

| <b>CAN Provisions</b>           | <b>Scope</b>                          |
|---------------------------------|---------------------------------------|
| CAN/CSA-C22.2<br>No. 61010-1    | Safety general laboratory equipment   |
| CAN/CSA C22.2<br>No. 61010-2-81 | Safety automatic laboratory equipment |

| <b>CHN Provisions</b> | <b>Scope</b>                               |
|-----------------------|--|
| Order 32/2016         | Restriction of hazardous substances (RoHS) |
| SJ/T 11364-2014       | Restriction of hazardous substances (RoHS) |

| <b>JPN Provisions</b> | <b>Scope</b>                                 |
|-----------------------|--|
| PSE (Denan) Law       | Electrical appliance and material safety law |

| <b>EAC Технический регламент Таможенного союза</b> |  |
|--|--|
| TP TC 004/2011                                     | О безопасности низковольтного оборудования         |
| TP TC 020/2011                                     | Электromагнитная совместимость технических средств |

| <b>KOR Provisions</b> |   |
|-----------------------|---|
| KC 61010-1            | Safety general laboratory equipment     |
| KC 61010-2-081        | Safety automatic laboratory equipment   |
| KC 62368-1            | Safety information technology equipment |

| <b>AUS/NZL Provisions</b> |   |
|---------------------------|---|
| AS/NZ 61010-1:2003        | Safety general laboratory equipment     |
| IEC 61010-2-081:2019      | Safety automatic laboratory equipment   |
| AS/NZ 62368-1:2022        | Safety information technology equipment |

### **RoHS compliance statement**

This product complies with the requirements of Directive 2011/65/EU (RoHS), applying the following exemptions according to Annex III:

- 7(c)-I – Electrical and electronic components containing lead in glass or ceramic (except dielectric ceramic in capacitors).
- 7(a) – Lead in high melting point solders (lead alloys with a lead content of more than 85 wt%).
- 6(a)-I – Lead as an alloying element in steel for machining purposes (up to 0.35 % lead) and in galvanized steel components (up to 0.2 % lead).

Currently valid; an extension until 31 December 2026 has been announced but is not yet legally in force.

**Additional information according to REACH (Article 33)**

Electronic components of this product contain the following substances in a concentration above 0.1 % w/w:

- Lead (CAS 7439-92-1)
- 1,2-Dimethoxyethane (CAS 110-71-4)
- 2,2'-Methylenebis[4-methyl-6-tert-butylphenol] (CAS 119-47-1)

Please follow standard precautions when handling.

**California Proposition 65 notice**

This product contains chemicals such as lead, nickel, lead silicon oxide (PbSiO<sub>2</sub>), nickelous oxide, carbon black (airborne, unbound particles of respirable size), cobalt, and antimony oxide, which are listed by the State of California under Proposition 65. Additional information about Proposition 65 is available at [www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov).

The design of this device is fully enclosed, and all electronic components containing these substances are completely integrated and not accessible during normal use. As a result, user exposure to these chemicals is not expected under typical operation.

Zizers, 2026-03-12



Urs Hartmann  
CEO



Daniela Gross  
Head of Corporate Quality