

Analytical data for the routine testing of GRIPTIPS® and CLEAR ADVANTAGE® reservoirs:

PROPERTY	TEST METHOD	LIMIT OF DETECTION	RESULT
Human DNA	Internal test method	< 30 pg DNA	Pass
DNase	Internal test method	< 1 x 10E-6 Kunitz units	Pass
RNase	Internal test method	< 1 x 10E-9 Kunitz units	Pass
PCR inhibition	Internal test method	No difference at 40 cycles endpoint	Pass
Endotoxin	USP <85>	< 0.06 EU/ml	Pass

Above defined items are regularly tested by the manufacturer through ISO 17025-certified reference laboratories. Our products are manufactured, packaged, stored, and shipped in compliance with ISO 9001-certified processes. INTEGRA products are produced using precise and consistent manufacturing techniques to ensure excellent performance and quality. We strive to exceed customer expectations and actively encourage continuous feedback. Our manufacturing process includes consistent and thorough inspections at every stage, ensuring that products remain contamination-free and consistent across all lots.

INTEGRA Biosciences AG



Daniela Gross, Global Head of Quality

06. Feb. 2025

Date