

# Safety regulations of automated laboratory equipment

**INTEGRA Biosciences** is committed to provide reliable, high-quality products which are safe and environmental friendly – Compliance with relevant safety legislation worldwide is a key part of **INTEGRA Biosciences'** mission. This document covers regulatory aspects of product safety of automated laboratory equipment produced by **INTEGRA Biosciences**.

Examples of such equipment include **ASSIST** (Part No 4500), **ASSIST PLUS** (Part No 4505), **VIAFLO 96** (Part No 6001), **VIAFLO 384** (Part No 6031), **WELLJET** (Part No 500x), **MINI 96** (Part No 480x) and **MEDIAJET** (Part No 1x300x).

**INTEGRA Biosciences** continuously monitors changes in global product safety regulations to ensure ongoing compliance. This document will be updated as necessary to reflect any relevant changes.

## 1 Applicable standards

The IEC 61010 standard series cover all aspects of laboratory equipment safety, e.g. mechanical aspects, sound emissions, laser sources, pressure, fire and ergonomic aspects. This standard series is recognized worldwide, such as UL 61010 in the **United States of America (USA)** or EN 61010 in the **European Union (EU)**. This standard series also covers other **INTEGRA Biosciences** instruments including **MEDIACLAVE**, **FIREBOY**, **VACUSAFE**, **PIPETBOY** etc.

Part 1 of this standard series describes general requirements whereas Part 2 is divided into multiple sections that address specific safety aspects of equipment, e.g. automatic laboratory equipment for analysis (-2-81). The scope of IEC 61010-2-81 is defined as: "instruments or systems for measuring or modifying one or more characteristics or parameters of samples, performing the complete process or parts of the process without manual intervention". The **INTEGRA Biosciences** instruments mentioned in the introduction are thus clearly in the scope of both IEC 61010-1 and IEC 61010-2-81.

## 2 Risk mitigation

All **INTEGRA Biosciences** instruments are carefully designed with a strong focus on handling safety. A thorough risk analysis is performed with **Failure Mode and Effects Analysis (FMEA)** and methods described in IEC 61010 section 7 and Annex J covering all operational aspects and the entire product life cycle.

Limited speed and force range combined with sensors to detect blockage or user interference guarantee the lowest possible risk (category 1, "broadly acceptable") as defined by IEC 61010. Additional measures such as warning markings, audible and visual signals and instructions for use allow to safely interact with the equipment without the need for protective hoods or cabinets. All in-house and independent TÜV SÜD Product Services safety tests have been successfully passed.

### 3 Legal considerations EU

By **EU** law conformity can be declared to either the **Low Voltage Directive (LVD)** or the **Machinery Directive (MD)**.

The current definition of “machinery” according to **MD** Article 2a is “an assembly, fitted with or intended to be fitted with a drive system other than directly applied human or animal effort, consisting of linked parts or components, at least one of which moves, and which are joined together for a specific application”. This scope covers almost any type of equipment with moving parts powered by stored energy, which in many cases (e.g. “ball pen”) was never intended to fall under the **MD**.

However, regarding other directives article 3 of the **MD** states “Where, for machinery, the hazards referred to in Annex I are wholly or partly covered more specifically by other Community Directives, this Directive shall not apply, or shall cease to apply, to that machinery in respect of such hazards from the date of implementation of those other Directives”.

General laboratory equipment is covered in scope by standard EN 61010-1, the European version of IEC 61010-1. This standard is harmonized under the **LVD** (see [https://ec.europa.eu/growth/single-market/european-standards/harmonised-standards/low-voltage\\_en](https://ec.europa.eu/growth/single-market/european-standards/harmonised-standards/low-voltage_en)).

EN 61010-2-81 explicitly lists “automatic sampler / pipettor / aliquoter” in its extended scope definition. Its general scope is defined as “automatic and semi-automatic laboratory equipment for analysis and other purposes”. Therefore, this standard specifically applies to instruments like **ASSIST, ASSIST PLUS, VIAFLO 96, VIAFLO 384, MINI 96, WELLJET** and **MEDIAJET**. EN 61010-2-81 is also harmonized under the **LVD**.

What’s more, to **INTEGRA Biosciences** knowledge there is no court decision setting a legal precedent favoring **MD** over **LVD** for aforementioned type of equipment.

Based on the above **INTEGRA Biosciences** laboratory instruments are most specifically covered by requirements of the IEC 61010 series of standards harmonized under the **LVD**. Conformity is therefore declared to the **LVD**, not the **MD**. This opinion is shared by a majority of manufacturers of similar equipment.

### 4 Signatures

Zizers, August 2025



Urs Hartmann  
CEO INTEGRA Biosciences AG



Daniela Gross  
Head of Corporate Quality

If you require further information or a more detailed explanation on a specific aspect, please do not hesitate to contact us!

**INTEGRA Biosciences AG**  
7205 Zizers, Switzerland  
T +41 81 286 95 30  
info@integra-biosciences.com

**INTEGRA Biosciences Corp.**  
Hudson NH 03051, USA  
T +1 603 578 5800  
info-us@integra-biosciences.com

**INTEGRA Biosciences Deutschland GmbH**  
35444 Biebertal, Deutschland  
T +49 6409 81 999 15  
info-de@integra-biosciences.com

**INTEGRA Biosciences SAS**  
95062 Cergy-Pontoise Cedex 1, France  
T +33 1 34 30 76 76  
info-fr@integra-biosciences.com

**INTEGRA Biosciences Ltd.**  
Thatcham Berkshire RG194EP, UK  
T +44 1635 797 00  
info-uk@integra-biosciences.com

**INTEGRA Biosciences (Shanghai) Co., Ltd.**  
上海自由贸易试验区 201315, 中国  
T +86 21 5844 7203  
info-cn@integra-biosciences.com

**INTEGRA Biosciences KK**  
ちよだ一く、ときよ 101-0031、じゃぱん  
T +81 3 5962 4936  
info-jp@integra-biosciences.com

**INTEGRA Biosciences Nordic ApS**  
2605 Brøndby, Denmark  
T +45 3173 5373  
orders-nordic@integra-biosciences.com

**INTEGRA Biosciences Ireland Limited**  
Blanchardstown, Dublin 15, Ireland  
T +44 1635 797000  
info-ie@integra-biosciences.com

**INTEGRA Biosciences Pty Ltd**  
Brendale QLD 4500, Australia  
T +61 7 3497 5800  
info-au@integra-biosciences.com

**INTEGRA Biosciences Benelux B.V**  
4814 DB Breda, Netherlands  
T +32 78 48 66 75  
orders-benelux@integra-biosciences.com