



CERTIFICATE

No. U8 042035 0013 Rev. 00

Holder of Certificate: **INTEGRA Biosciences AG**

Tardisstrasse 201
7205 Zizers
SWITZERLAND

Certification Mark:



C US

Product: **Laboratory Equipment
(media preparation)**

Model(s): **MEDIACLAVE 10**

Parameters: Rated voltage: 200-240V
Rated frequency: 50/60Hz
Rated power : 3750W
Protection class: I

Tested according to: UL 61010-1:2012/R:2024-11
UL 61010-2-010:2019/R:2023-09
CSA C22.2 No. 61010-1:2012/U4:2024-11
CSA C22.2 No. 61010-2-010:2019

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing, Certification, Validation and Verification Regulations (TCVVR)". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713405686-00

Date, 2026-06-02

Siemon

(Thorsten Siemon)