

Channel	Part No.	Volume Range (µl)	Volume Increments (µl)	Test Volume (µl)	Accuracy (±%)	Precision (≤%)
1	4011	0.5 - 12.5	0.01	1.25	5.00	4.00
				6.25	1.50	0.80
				12.5	1.00	0.40
	4016	2 - 50	0.05	5	3.00	1.50
				25	1.50	0.60
				50	1.00	0.40
4012	5 - 125	0.1	12.5	3.00	1.00	
			62.5	1.20	0.40	
			125	0.60	0.20	
4013	10 - 300	0.5	30	2.00	0.60	
			150	1.00	0.30	
			300	0.60	0.15	
4014	50 - 1250	1	125	3.00	0.60	
			625	1.00	0.30	
			1250	0.60	0.17	
4015	100 - 5000	5	500	3.00	0.75	
			2500	1.20	0.30	
			5000	0.60	0.15	
8	4621	0.5 - 12.5	0.01	1.25	10.00	6.00
				6.25	4.00	1.60
				12.5	2.00	0.80
	4626	2 - 50	0.05	5	5.00	2.50
				25	2.50	0.60
				50	1.50	0.40
4622	5 - 125	0.1	12.5	3.75	1.50	
			62.5	2.50	0.70	
			125	1.60	0.35	
4623	10 - 300	0.5	30	4.00	1.20	
			150	2.00	0.60	
			300	1.60	0.35	
4624	50 - 1250	1	125	6.00	1.10	
			625	2.40	0.50	
			1250	1.20	0.30	
12	4631	0.5 - 12.5	0.01	1.25	10.00	6.00
				6.25	4.00	1.60
				12.5	2.00	0.80
	4636	2 - 50	0.05	5	5.00	2.50
				25	2.50	0.60
				50	1.50	0.40
4632	5 - 125	0.1	12.5	3.75	1.50	
			62.5	2.50	0.70	
			125	1.60	0.35	
4633	10 - 300	0.5	30	4.00	1.20	
			150	2.00	0.60	
			300	1.60	0.35	
4634	50 - 1250	1	125	6.00	1.10	
			625	2.40	0.50	
			1250	1.20	0.30	
16	4641	0.5 - 12.5	0.01	1.25	10.00	6.00
				6.25	4.00	1.60
				12.5	2.00	0.80
	4646	2 - 50	0.05	5	5.00	2.50
				25	2.50	0.60
				50	1.50	0.40
4642	5 - 125	0.1	12.5	3.75	1.50	
			62.5	2.50	0.70	
			125	1.60	0.35	

1. The multichannel specifications represent the performance of all channels of a pipette.

2. Precision = coefficient of variation