

| Kanal | Art.-Nr. | Volumen-Bereich (µl) | Volumen-schritt (µl) | Test-Volumen (µl) | Genauigkeit (±%) | Präzision (≤%) |
|-------|----------|----------------------|----------------------|-------------------|------------------|----------------|
| 24 | 6123 | 10 – 300 | 0.5 | 30 | 2.50 | 1.50 |
| | | | | 150 | 1.50 | 1.00 |
| | | | | 300 | 1.20 | 0.80 |
| | 6124 | 50 – 1250 | 1 | 125 | 2.50 | 1.50 |
| | | | | 625 | 2.00 | 1.00 |
| | | | | 1250 | 1.80 | 0.80 |
| 96 | 6101 | 0.5 – 12.5 | 0.01 | 1.25 | 8.00 | 3.00 |
| | | | | 6.25 | 2.00 | 1.50 |
| | | | | 12.5 | 1.50 | 1.00 |
| | 6106 | 2 – 50 | 0.05 | 5 | 5.00 | 2.50 |
| | | | | 25 | 2.00 | 1.50 |
| | | | | 50 | 1.50 | 1.00 |
| | 6102 | 5 – 125 | 0.1 | 12.5 | 3.00 | 1.70 |
| | | | | 62.5 | 2.00 | 1.50 |
| | | | | 125 | 1.50 | 1.00 |
| | 6103 | 10 – 300 | 0.5 | 30 | 2.50 | 1.50 |
| | | | | 150 | 1.50 | 1.00 |
| | | | | 300 | 1.20 | 0.80 |
| | 6104 | 50 – 1250 | 1 | 125 | 2.50 | 1.50 |
| | | | | 625 | 2.00 | 1.00 |
| | | | | 1250 | 1.80 | 0.80 |
| 384 | 6131 | 0.5 – 12.5 | 0.01 | 1.25 | 8.00 | 3.00 |
| | | | | 6.25 | 2.00 | 1.50 |
| | | | | 12.5 | 1.50 | 1.00 |
| | 6136 | 2 – 50 | 0.05 | 5 | 5.00 | 2.50 |
| | | | | 25 | 2.00 | 1.50 |
| | | | | 50 | 1.50 | 1.00 |
| 6132 | 5 – 125 | 0.1 | 12.5 | 3.00 | 1.70 | |
| | | | 62.5 | 2.00 | 1.50 | |
| | | | 125 | 1.50 | 1.00 | |

1. Die Spezifikationen bilden die Eigenschaften aller Pipettierkanäle ab.
2. Spezifikationen gelten nur für Einzelpipettierungen im Modus "Pipettieren".
3. Präzision = Variationskoeffizient